

ATP Side By Side Evaluation Kit Instructions

Description/Intended Use

Adenosine triphosphate (ATP) hygiene monitoring systems detect levels of ATP from both microbial and non-microbial contamination. The amount of ATP collected and measured in systems is expressed in terms of relative light units (RLU). Variation in results between systems can be caused by the enzyme reagent formulation used to produce the bioluminescent reactions, the extractant pre-moistened on the swab bud, the electronic calibration within the luminometer, and/or the variation in the sample being collected. Understanding the correlation of ATP levels to RLU is important when comparing systems. Simple surface sampling comparisons can be highly variable due to sampling technique, surface type, sample type, and possible extreme variations in residue present in different areas on the same surface.

The Side by Side Evaluation Kit eliminates sampling error and provides a consistent and scientifically-based method for comparing systems by pipetting a known amount of ATP directly onto the tips of testing devices. The instruction sheet describes the procedure for comparing two ATP monitoring systems. A monitoring system is defined as any combination of luminometer and compatible ATP test device

Provided Materials

- 25 Hygiene test devices
- 3 vials ATP dilutions (2 nM, 20 nM, and 200 nM)
- (1) 10uL pipette
- 1 pair sterile gloves
- 1 data record sheet
- Download the [ATP Side by Side Evaluation Worksheet](#)

Required Materials (not provided in the kit):

- Luminometer(s) for comparison
- (25) ATP test devices comparison

Instructional Video

ATP Standards

1. Allow ATP vials and ATP test devices to equilibrate to room temperature (10 minutes at 21 - 25 C) before use.
2. Using an aseptic technique, carefully remove caps from vials.
3. Turn on luminometer(s)
4. Remove 10uL pipette from the bag. Leave pipette tips in a bag or place them where they will not be contaminated. Place on the pipette tip on the end of the pipette. Be careful not to touch the top of the pipette tip as this could contaminate the ATP standards.
5. Place the pipette 10uL of 2nM ATP standard directly onto the top of one ATP test device from the first set of 25.
6. Activate and measure in the instrument according to instructions.
7. Record result on data record sheet or input directly into Excel worksheet
8. Repeat steps 5 - 7 for more times to give a total of 5 replicates. Use a new pipette tip for each aliquot sample tested.
9. Repeat steps 5 - 8 with 20 nM ATP.
10. Repeat steps 5 - 8 with 200 nM ATP.
11. Repeat steps 5 - 10 with other monitoring systems.

Background

1. Background is determined by testing blank swabs (i.e., without any added sample). Without opening the device, activate and measure 10 test devices for each monitoring system.
2. Record results for each.

Interpretation of Results

[ATP Side by Side Evaluation Worksheet](#) performs all calculations automatically. When comparing results, consider background, repeatability, linearity, sensitivity, and Pass/Fail correlation.

Safety & Precautions

Components of Hygiena ATP test devices do not pose any health risk when used in accordance with standard laboratory practice and procedures of this insert

- Test devices are for one-time use. Do not reuse it. For further safety instruction, refer to Safety Data Sheet (SDS).

Hygiena Liability

Hygiena will not be liable to User or others for any loss or damage whether direct or indirect, incidental, or consequential from use of this device. If this product is proven to be defective, Hygiena's sole obligation will be to replace product or at its discretion, refund the purchase price. Promptly notify Hygiena within 5 days of discovery of any suspected defect and return the product to Hygiena. Please contact Customer Service for a Returned Goods Authorization number.

Contact Information

If more information is required please visit us at www.hygiena.com

Contact Hygiena Technical Support for further assistance.

- Phone: 1-888-HYGIENA (1-888-494-4362, option 2)
- Email: techsupport@hygiena.com
- [Submit a Support Ticket](#)
- [Schedule a Microsoft Teams meeting with support](#)